

FCC Supplier Declaration of Conformity

Pico Sales and Distribution as the official importer hereby declares that the products listed below meet the all the essential requirements as set out in FCC regulations 47 CFR part 15 subpart B as required by 47 CFR section 2.906. Technical documentation required to demonstrate compliance to the standards is available for inspection by the relevant enforcement authorities.

Products covered by this declaration:

PicoScope 4444 4 channel high-resolution differential USB oscilloscope.

The basis on which conformity is being declared:

CFR 47	Code of Federal Regulations FCC: part 15 Subpart B - Frequency devices - unintentional

Radiators. Radiated emissions standard. Class A emissions.

ANSI C63.4:2014 American National Standard for Methods of Measurement of Radio-Noise Emissions from Low-

Voltage Electrical and Electronic Equipment in the Range of 9 kHz to 40 GHz

EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements. Group 1,

Class A equipment. (Emissions section only)

EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements. Basic

Environment. (Immunity section only)

EN 61326-2-1:2013 Part 2-1: Test configurations, operational conditions and performance criteria for sensitive test and

measurement equipment for unprotected applications.

Electromagnetic compatibility (EMC) - Part 3-3: limits - Limitation of voltage changes, voltage EN 61000-3-3:2013

fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤

16 A per phase and not subject to conditional connection.

EN 61000-3-2:2014 Electromagnetic compatibility (EMC) - Part 3-2 Limits - Limits for harmonic current emissions for

equipment input current up to and including 16 A per phase.

Craig Wheeler

VP Finance & Admin

Signed: Mark All Date: 2 Hoch / 2012

Location:

Tyler TX