

## UK Declaration of Conformity

This Declaration of Conformity is issued under the sole responsibility of Pico Technology and declares that the following products comply with the requirements of the specified Statutory Requirements and Standards as listed below. Technical documentation required to demonstrate compliance to the standards is available for inspection by the relevant enforcement authorities.

*Products covered by this declaration:*

<b>PicoScope 3203D</b>	<b>2-channel 50MHz USB oscilloscope</b>
<b>PicoScope 3204D</b>	<b>2-channel 70MHz USB oscilloscope</b>
<b>PicoScope 3205D</b>	<b>2-channel 100MHz USB oscilloscope</b>
<b>PicoScope 3206D</b>	<b>2-channel 200MHz USB oscilloscope</b>
<b>PicoScope 3403D</b>	<b>4-channel 50Mhz USB oscilloscope</b>
<b>PicoScope 3404D</b>	<b>4-channel 70Mhz USB oscilloscope</b>
<b>PicoScope 3405D</b>	<b>4-channel 100Mhz USB oscilloscope</b>
<b>PicoScope 3406D</b>	<b>4-channel 200Mhz USB oscilloscope</b>

*The object of the declaration is in conformity with the relevant Statutory Requirements:*

SI 2016/1101	Electrical Equipment (Safety) Regulations
SI 2016/1091	EMC Regulations
SI 2012/3032	Restriction of Hazardous Substances Regulations
SI 2013/3113	Waste Electrical and Electronic Equipment Regulations

*The basis on which conformity is being declared:*

EN 61010-1:2010+A1:2019	Safety requirements for electrical equipment for measurement, control and laboratory use, general equipment requirements.
EN 61010-2-030:2010	Particular requirements for testing and measuring circuits.
EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use - EMC requirements. Group 1, Class A equipment. (Emissions section only)
EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use - EMC requirements. Basic Environment. (Immunity section only)
EN 61326-2-1:2013	Part 2-1: Test configurations, operational conditions and performance criteria for sensitive test and measurement equipment for unprotected applications.
EN 61000-3-3:2013	Electromagnetic compatibility (EMC) - Part 3-3: limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection.

All component parts are supplied to Pico Technology under a valid Declaration of Conformity declaring their RoHS compliance. Additional analysis may be performed and recorded by Pico Technology Ltd or its subcontractors of the materials used in the manufacture of the above products as required by the Pico Technology quality standards.

Roger Widdowson  
Compliance Manager

Signed: 

Date: 26<sup>th</sup> Sept 2022



Pico Technology  
James House  
Colmworth Business Park  
St Neots PE19 8YP  
United Kingdom

www.picotech.com  
www.picoauto.com  
contact@picotech.com

T: +44 (0) 1480 396 395



FM52022 EMS612349  
Registered in England & Wales No: 2626181