



Pico Technology
James House, Marlborough Road.
Colmworth Business Park.
Eaton Socon, St Neots, Cambridgeshire.
PE19 8YP United Kingdom.
Tel: +44 1480 396395. Fax: +44 1480 396296

EC Declaration of Conformity

Pico Technology declares that the following product complies with the requirements of the specified Directives and Standards as listed below. Technical documentation required to demonstrate compliance to the standards is available for inspection by the relevant enforcement authorities. Product carries the CE mark.

Products covered by this Declaration:

PicoLog USB DrDAQ multi function data logger.

EU Directives covered by this Declaration:

2004/108/EC Electromagnetic Compatibility Directive.
2006/95/EC Low Voltage Equipment Directive.

The Basis on which Conformity is being Declared:

EN61010-1:2001	Safety requirements for electrical equipment for measurement, control and laboratory use, general equipment requirements.
EN61326-1:2006	EMC Immunity and Emissions for measurement, control and laboratory equipment - general requirements. Test limits and frequencies are specified in CISPR11 and EN61000-4.
CISPR11:2006	Industrial, scientific and medical equipment – radio frequency disturbance characteristics – limits and methods of measurement. Radiated and Conducted emissions. Class A emissions.
CFR 47:2009	Code of Federal Regulations FCC: part 15 Subpart B – Frequency devices – unintentional Radiators. Radiated emissions standard. Class A emissions.
EN61000-4 EN61000-4-2:1995 +A1:1998 +A2:2001	Radiated and Conducted Immunity., including Electrostatic Discharge.
EN61000-4-3:2006	Radiated RF.
EN61000-4-4:2004	Electrical fast transients and bursts.
EN61000-4-6:2007	Conducted RF

Alan Tong
Managing Director
Dec 2010

Signed

A handwritten signature in black ink, consisting of several fluid, overlapping strokes.